

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JAN 13 2011 2011  
3/9/11

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130117</u> ✓	
Name <u>KYLE VALERIO</u>	ID# <u>2093</u> ✓ Date <u>1/10/2011</u> ✓
A Agency <u>KODIAK POLICE DEPARTMENT</u> Phone # <u>(907)486-8000</u>	
Instrument Location <u>217 LOWER MILL BAY ROAD, KODIAK, AK. 99615</u>	
B Alco S/N <u>X172961</u> ✓ Target Value <u>.081</u> ✓ High Pressure <u>1100</u>	
Alco Test Values <u>.079</u> ✓ <u>.080</u> ✓ 1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 11<sup>th</sup> day of MAR, 2011

(Notary Seal Stamp)

Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm breath Test Instrument*

*State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130117

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117

JANUARY 10, 2011

OPERATOR'S NAME: *amb 2/8/11*  
VALERIO, KYLE D  
OPERATOR'S NUMBER: 2893  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
NEW ALCO

O.L. #: XXXXXX  
DEPT/AGENCY: KOD1  
CASE/REPORT: 2011000000  
TEST TYPE: V  
ALCO TARGET VALUE: .081  
ALCO S/N: K172961

### — BREATH ANALYSIS —

.081 ADJUSTED FOR 30.77 in		
ALCO TARGET	.083	12:07
BLANK TEST	.000	12:08
INTERNAL STANDARD	VERIFIED	12:08
ALCO TU 30.77 in	.079	12:08
BLANK TEST	.000	12:09
SUBJECT SAMPLE	.000	12:10
BLANK TEST	.000	12:11
ALCO TU 30.77 in	.000	12:11
BLANK TEST	.000	12:12

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117

JANUARY 10, 2011  
TIME 12:14

### — DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.77 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~■